

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX042993.D
 Acq On : 09 Aug 2024 13:53
 Operator : JC/MD
 Sample : P3514-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR-05-459-475-080724

Quant Time: Aug 09 16:27:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	158818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	260320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115073	52.780	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.560%		
35) Dibromofluoromethane	5.391	113	85051	49.968	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.940%		
50) Toluene-d8	8.652	98	320609	53.059	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.120%		
62) 4-Bromofluorobenzene	11.079	95	106036	48.834	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 97.660%		
Target Compounds						Qvalue
16) Acetone	2.391	43	6287	6.648	ug/l	94
52) Toluene	8.720	92	4516	1.039	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX042993.D
Acq On : 09 Aug 2024 13:53
Operator : JC/MD
Sample : P3514-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR-05-459-475-080724

Quant Time: Aug 09 16:27:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

