

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX043000.D
 Acq On : 09 Aug 2024 17:01
 Operator : JC/MD
 Sample : P3430-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 927-K1-WPO-0-0.25-073124-FD

Quant Time: Aug 09 23:27:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240632	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104312	50.495	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.000%	
35) Dibromofluoromethane	5.385	113	81406	51.740	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.480%	
50) Toluene-d8	8.647	98	298341	53.413	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.820%	
62) 4-Bromofluorobenzene	11.079	95	97940	48.796	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.600%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	14622	8.720	ug/l	92
16) Acetone	2.392	43	8184	9.134	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	1819	1.090	ug/l #	54
27) cis-1,2-Dichloroethene	4.495	96	5631	2.845	ug/l	81
29) Tetrahydrofuran	5.038	42	6216	6.537	ug/l	94
52) Toluene	8.720	92	8046	2.004	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX043000.D
Acq On : 09 Aug 2024 17:01
Operator : JC/MD
Sample : P3430-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
927-K1-WPO-0-0.25-073124-FD

Quant Time: Aug 09 23:27:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

