

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX043002.D
 Acq On : 09 Aug 2024 17:48
 Operator : JC/MD
 Sample : P3457-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 924-K1-WS-080224

Quant Time: Aug 09 23:28:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148730	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	240279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218815	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106806	52.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.620%	
35) Dibromofluoromethane	5.385	113	79856	50.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.660%	
50) Toluene-d8	8.647	98	298083	53.446	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.900%	
62) 4-Bromofluorobenzene	11.079	95	100010	49.901	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.800%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	4340	4.901	ug/l	# 76
52) Toluene	8.726	92	1342	0.335	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX043002.D
Acq On : 09 Aug 2024 17:48
Operator : JC/MD
Sample : P3457-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
924-K1-WS-080224

Quant Time: Aug 09 23:28:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

