

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX043003.D
 Acq On : 09 Aug 2024 18:11
 Operator : JC/MD
 Sample : P3430-05DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 925-K1-WPO-0-0.25-073124DL

Quant Time: Aug 10 00:27:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	164547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	269486	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117927	52.205	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.420%		
35) Dibromofluoromethane	5.385	113	89756	50.939	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.880%		
50) Toluene-d8	8.647	98	334112	53.413	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.820%		
62) 4-Bromofluorobenzene	11.079	95	110687	49.242	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.480%		
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	9555	5.211	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.319	101	644	0.359	ug/l #	74
27) cis-1,2-Dichloroethene	4.489	96	91111	42.104	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX043003.D
Acq On : 09 Aug 2024 18:11
Operator : JC/MD
Sample : P3430-05DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
925-K1-WPO-0-0.25-073124DL

Quant Time: Aug 10 00:27:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

