

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003973.D
 Acq On : 10 Aug 2018 13:37
 Operator : JC/MD
 Sample : J4379-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-103I

Quant Time: Aug 13 10:46:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	302700	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	478222	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	435198	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	237839	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	212270	52.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.80%	
35) Dibromofluoromethane	5.51	113	185319	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	8.72	98	669701	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
62) 4-Bromofluorobenzene	11.14	95	228327	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
Target Compounds						
24) 1,1-Dichloroethane	3.70	63	7963	1.16	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	6272	1.57	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003973.D
Acq On : 10 Aug 2018 13:37
Operator : JC/MD
Sample : J4379-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-103I

Quant Time: Aug 13 10:46:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

