

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003974.D
 Acq On : 10 Aug 2018 14:04
 Operator : JC/MD
 Sample : J4379-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102S

Quant Time: Aug 13 10:47:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	302099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	479159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	450958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	248579	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	214432	53.55	ug/l	0.00
Spiked Amount						
				Recovery	=	107.10%
35) Dibromofluoromethane	5.51	113	184220	50.75	ug/l	0.00
Spiked Amount						
				Recovery	=	101.50%
50) Toluene-d8	8.72	98	674751	47.80	ug/l	0.00
Spiked Amount						
				Recovery	=	95.60%
62) 4-Bromofluorobenzene	11.14	95	233780	51.18	ug/l	0.00
Spiked Amount						
				Recovery	=	102.36%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	9.34	164	2397	0.54	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003974.D
Acq On : 10 Aug 2018 14:04
Operator : JC/MD
Sample : J4379-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102S

Quant Time: Aug 13 10:47:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

