

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003977.D
 Acq On : 10 Aug 2018 15:24
 Operator : JC/MD
 Sample : J4379-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101S

Quant Time: Aug 13 10:49:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	317734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	501383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	466653	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	249803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	221842	52.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.34%	
35) Dibromofluoromethane	5.51	113	192754	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
50) Toluene-d8	8.72	98	699361	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
62) 4-Bromofluorobenzene	11.14	95	243416	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	14152	4.15	ug/l	94
24) 1,1-Dichloroethane	3.70	63	22452	3.13	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	13551	3.23	ug/l	94
30) Chloroform	5.22	83	10006	1.48	ug/l	89
44) Trichloroethene	7.21	130	124001	26.89	ug/l	96
64) Tetrachloroethene	9.34	164	49774	10.84	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003977.D
Acq On : 10 Aug 2018 15:24
Operator : JC/MD
Sample : J4379-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101S

Quant Time: Aug 13 10:49:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

