

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003978.D
 Acq On : 10 Aug 2018 15:50
 Operator : JC/MD
 Sample : J4379-23
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101I

Quant Time: Aug 13 10:50:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	437702	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	415618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217909	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	198467	53.69	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.38%	
35) Dibromofluoromethane	5.51	113	171282	51.65	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.30%	
50) Toluene-d8	8.72	98	613317	47.56	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.12%	
62) 4-Bromofluorobenzene	11.14	95	210762	50.51	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	2.37	96	11838	3.96	ug/l	96
24) 1,1-Dichloroethane	3.70	63	13299	2.11	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	10260	2.78	ug/l	96
30) Chloroform	5.21	83	3342	0.56	ug/l	93
44) Trichloroethene	7.21	130	49983	12.42	ug/l	92
64) Tetrachloroethene	9.34	164	118149	28.90	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003978.D
Acq On : 10 Aug 2018 15:50
Operator : JC/MD
Sample : J4379-23
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101I

Quant Time: Aug 13 10:50:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

