

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003980.D
 Acq On : 10 Aug 2018 16:43
 Operator : JC/MD
 Sample : J4379-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-103D

Quant Time: Aug 13 10:56:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	300698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	461590	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	428237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	226563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	210358	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	5.51	113	181316	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
50) Toluene-d8	8.72	98	637899	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
62) 4-Bromofluorobenzene	11.14	95	219125	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	2347	0.73	ug/l	97
24) 1,1-Dichloroethane	3.70	63	8089	1.19	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	4674	1.18	ug/l	83
44) Trichloroethene	7.21	130	3054	0.72	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003980.D
Acq On : 10 Aug 2018 16:43
Operator : JC/MD
Sample : J4379-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-103D

Quant Time: Aug 13 10:56:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

