

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003981.D
 Acq On : 10 Aug 2018 17:10
 Operator : JC/MD
 Sample : J4397-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Aug 13 10:56:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	303699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	463299	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	424532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	233205	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207938	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	5.51	113	180520	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
50) Toluene-d8	8.72	98	644726	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
62) 4-Bromofluorobenzene	11.14	95	221343	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
Target Compounds						
39) Methylcyclohexane	7.46	83	1513	0.25	ug/l	# 90
44) Trichloroethene	7.21	130	2042	0.48	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003981.D
Acq On : 10 Aug 2018 17:10
Operator : JC/MD
Sample : J4397-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Aug 13 10:56:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

