

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003982.D
 Acq On : 10 Aug 2018 17:37
 Operator : JC/MD
 Sample : J4397-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Aug 13 10:57:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	281365	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	449919	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	413444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216159	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	206297	55.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.62%	
35) Dibromofluoromethane	5.51	113	169736	49.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.60%	
50) Toluene-d8	8.72	98	628822	47.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.88%	
62) 4-Bromofluorobenzene	11.14	95	207241	48.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003982.D
Acq On : 10 Aug 2018 17:37
Operator : JC/MD
Sample : J4397-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
FIELD-BLANK

Quant Time: Aug 13 10:57:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

