

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003984.D
 Acq On : 10 Aug 2018 18:30
 Operator : JC/MD
 Sample : J4379-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-08082018

Quant Time: Aug 13 10:58:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	478105	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	442801	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	237772	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	218244	52.48	ug/l	0.00
Spiked Amount			Recovery	=	104.96%	
35) Dibromofluoromethane	5.51	113	182705	50.44	ug/l	0.00
Spiked Amount			Recovery	=	100.88%	
50) Toluene-d8	8.72	98	659137	46.80	ug/l	0.00
Spiked Amount			Recovery	=	93.60%	
62) 4-Bromofluorobenzene	11.14	95	225929	49.57	ug/l	0.00
Spiked Amount			Recovery	=	99.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003984.D
Acq On : 10 Aug 2018 18:30
Operator : JC/MD
Sample : J4379-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-08082018

Quant Time: Aug 13 10:58:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

