

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081018\
 Data File : VX003987.D
 Acq On : 10 Aug 2018 19:50
 Operator : JC/MD
 Sample : J4397-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-13

Quant Time: Aug 13 11:00:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	292707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	444511	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	412224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222406	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	208911	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
35) Dibromofluoromethane	5.51	113	176112	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
50) Toluene-d8	8.72	98	620027	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
62) 4-Bromofluorobenzene	11.14	95	212560	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
24) 1,1-Dichloroethane	3.70	63	3375	0.51	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081018\
Data File : VX003987.D
Acq On : 10 Aug 2018 19:50
Operator : JC/MD
Sample : J4397-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-13

Quant Time: Aug 13 11:00:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

