

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023625.D
 Acq On : 10 Aug 2021 09:41
 Operator : JC/MD
 Sample : VX0810MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0810MBL01

Quant Time: Aug 11 02:57:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	168551	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	286234	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113012	53.751	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.500%	
35) Dibromofluoromethane	5.391	113	91876	51.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.680%	
50) Toluene-d8	8.653	98	347155	52.797	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.600%	
62) 4-Bromofluorobenzene	11.085	95	122055	54.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
Data File : VX023625.D
Acq On : 10 Aug 2021 09:41
Operator : JC/MD
Sample : VX0810MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0810MBL01

Quant Time: Aug 11 02:57:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

