

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023626.D
 Acq On : 10 Aug 2021 10:05
 Operator : JC/MD
 Sample : VX0810WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0810WBL01

Quant Time: Aug 11 02:57:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	172495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	296559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	117545	54.628	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.260%		
35) Dibromofluoromethane	5.397	113	95346	51.421	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.840%		
50) Toluene-d8	8.653	98	356091	52.271	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 104.540%		
62) 4-Bromofluorobenzene	11.085	95	122668	52.954	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.900%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
Data File : VX023626.D
Acq On : 10 Aug 2021 10:05
Operator : JC/MD
Sample : VX0810WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0810WBL01

Quant Time: Aug 11 02:57:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

