

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023624.D
 Acq On : 10 Aug 2021 08:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 11 02:56:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.460	0.425	7.6	89	0.00
3 P	Chloromethane	0.536	0.491	8.4	91	0.00
4 C	Vinyl Chloride	0.542	0.493	9.0#	89	0.00
5 T	Bromomethane	0.322	0.282	12.4	93	0.00
6 T	Chloroethane	0.333	0.314	5.7	91	0.00
7 T	Trichlorofluoromethane	0.874	0.811	7.2	91	0.00
8 T	Diethyl Ether	0.323	0.292	9.6	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.483	2.4	96	0.00
10 T	Methyl Iodide	0.620	0.635	-2.4	95	0.00
11 T	Tert butyl alcohol	0.142	0.120	15.5	83	0.00
12 CM	1,1-Dichloroethene	0.491	0.446	9.2#	90	0.00
13 T	Acrolein	0.076	0.061	19.7	86	0.00
14 T	Allyl chloride	0.893	0.836	6.4	92	0.00
15 T	Acrylonitrile	0.322	0.292	9.3	87	0.00
16 T	Acetone	0.306	0.369	-20.6	126	0.00
17 T	Carbon Disulfide	1.211	0.973	19.7	79	0.00
18 T	Methyl Acetate	0.727	0.689	5.2	94	0.00
19 T	Methyl tert-butyl Ether	1.708	1.661	2.8	94	0.00
20 T	Methylene Chloride	0.595	0.532	10.6	95	0.00
21 T	trans-1,2-Dichloroethene	0.522	0.480	8.0	91	0.00
22 T	Diisopropyl ether	1.763	1.716	2.7	94	0.00
23 T	Vinyl Acetate	1.556	1.513	2.8	92	0.00
24 P	1,1-Dichloroethane	0.969	0.931	3.9	93	0.00
25 T	2-Butanone	0.475	0.483	-1.7	99	0.00
26 T	2,2-Dichloropropane	0.793	0.810	-2.1	99	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.582	4.0	94	0.00
28 T	Bromochloromethane	0.452	0.466	-3.1	104	0.00
29 T	Tetrahydrofuran	0.311	0.277	10.9	86	0.00
30 C	Chloroform	0.992	0.976	1.6#	97	0.00
31 T	Cyclohexane	0.924	0.831	10.1	91	0.00
32 T	1,1,1-Trichloroethane	0.848	0.848	0.0	95	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.636	-1.9	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.313	0.320	-2.2	103	0.00
36 T	1,1-Dichloropropene	0.452	0.447	1.1	94	0.00
37 T	Ethyl Acetate	0.570	0.523	8.2	89	0.00
38 T	Carbon Tetrachloride	0.460	0.465	-1.1	95	0.00
39 T	Methylcyclohexane	0.571	0.556	2.6	92	0.00
40 TM	Benzene	1.341	1.305	2.7	92	0.00
41 T	Methacrylonitrile	0.301	0.292	3.0	93	0.00
42 TM	1,2-Dichloroethane	0.496	0.515	-3.8	99	0.00
43 T	Isopropyl Acetate	0.884	0.869	1.7	92	0.00
44 TM	Trichloroethene	0.368	0.365	0.8	94	0.00
45 C	1,2-Dichloropropane	0.355	0.352	0.8#	93	0.00
46 T	Dibromomethane	0.241	0.246	-2.1	95	0.00
47 T	Bromodichloromethane	0.434	0.462	-6.5	98	0.00
48 T	Methyl methacrylate	0.409	0.425	-3.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023624.D
 Acq On : 10 Aug 2021 08:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 11 02:56:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	86	0.00
50 S	Toluene-d8	1.149	1.190	-3.6	104	0.00
51 T	4-Methyl-2-Pentanone	0.541	0.558	-3.1	94	0.00
52 CM	Toluene	0.831	0.854	-2.8#	96	0.00
53 T	t-1,3-Dichloropropene	0.493	0.549	-11.4	99	0.00
54 T	cis-1,3-Dichloropropene	0.541	0.568	-5.0	95	0.00
55 T	1,1,2-Trichloroethane	0.336	0.365	-8.6	98	0.00
56 T	Ethyl methacrylate	0.539	0.589	-9.3	98	0.00
57 T	1,3-Dichloropropane	0.572	0.597	-4.4	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.286	0.294	-2.8	98	0.00
59 T	2-Hexanone	0.410	0.449	-9.5	99	0.00
60 T	Dibromochloromethane	0.330	0.375	-13.6	99	0.00
61 T	1,2-Dibromoethane	0.359	0.381	-6.1	97	0.00
62 S	4-Bromofluorobenzene	0.391	0.459	-17.4	118	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.398	0.383	3.8	97	0.00
65 PM	Chlorobenzene	1.023	1.024	-0.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.378	-4.7	99	0.00
67 C	Ethyl Benzene	1.775	1.813	-2.1#	100	0.00
68 T	m/p-Xylenes	0.683	0.693	-1.5	100	0.00
69 T	o-Xylene	0.677	0.683	-0.9	101	0.00
70 T	Styrene	1.072	1.147	-7.0	103	0.00
71 P	Bromoform	0.266	0.283	-6.4	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.773	3.904	-3.5	104	0.00
74 T	N-amyl acetate	1.721	1.721	0.0	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.229	1.172	4.6	99	0.00
76 T	1,2,3-Trichloropropane	1.100	1.109	-0.8	102	0.00
77 T	Bromobenzene	0.910	0.949	-4.3	105	0.00
78 T	n-propylbenzene	4.270	4.476	-4.8	103	0.00
79 T	2-Chlorotoluene	2.552	2.663	-4.3	106	0.00
80 T	1,3,5-Trimethylbenzene	3.073	3.193	-3.9	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.406	0.7	97	0.00
82 T	4-Chlorotoluene	2.887	3.027	-4.8	106	0.00
83 T	tert-Butylbenzene	3.014	3.141	-4.2	104	0.00
84 T	1,2,4-Trimethylbenzene	3.056	3.215	-5.2	105	0.00
85 T	sec-Butylbenzene	3.908	4.092	-4.7	106	0.00
86 T	p-Isopropyltoluene	3.156	3.395	-7.6	106	0.00
87 T	1,3-Dichlorobenzene	1.684	1.712	-1.7	105	0.00
88 T	1,4-Dichlorobenzene	1.709	1.675	2.0	104	0.00
89 T	n-Butylbenzene	2.857	3.031	-6.1	107	0.00
90 T	Hexachloroethane	0.509	0.543	-6.7	104	0.00
91 T	1,2-Dichlorobenzene	1.630	1.632	-0.1	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.266	0.255	4.1	93	0.00
93 T	1,2,4-Trichlorobenzene	1.024	1.063	-3.8	105	0.00
94 T	Hexachlorobutadiene	0.460	0.454	1.3	105	0.00
95 T	Naphthalene	3.522	3.563	-1.2	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
Data File : VX023624.D
Acq On : 10 Aug 2021 08:54
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 11 02:56:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.034	1.070	-3.5	104	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6