

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023638.D
 Acq On : 10 Aug 2021 17:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 11 03:02:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.460	0.433	5.9	81	0.00
3 P	Chloromethane	0.536	0.508	5.2	85	0.00
4 C	Vinyl Chloride	0.542	0.501	7.6#	81	0.00
5 T	Bromomethane	0.322	0.296	8.1	88	0.00
6 T	Chloroethane	0.333	0.324	2.7	85	0.00
7 T	Trichlorofluoromethane	0.874	0.848	3.0	86	0.00
8 T	Diethyl Ether	0.323	0.315	2.5	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.507	-2.4	90	0.00
10 T	Methyl Iodide	0.620	0.674	-8.7	91	0.00
11 T	Tert butyl alcohol	0.142	0.126	11.3	79	0.00
12 CM	1,1-Dichloroethene	0.491	0.468	4.7#	85	0.00
13 T	Acrolein	0.076	0.058	23.7	74	0.00
14 T	Allyl chloride	0.893	0.874	2.1	87	0.00
15 T	Acrylonitrile	0.322	0.310	3.7	83	0.00
16 T	Acetone	0.306	0.274	10.5	84	0.00
17 T	Carbon Disulfide	1.211	0.991	18.2	72	0.00
18 T	Methyl Acetate	0.727	0.729	-0.3	89	0.00
19 T	Methyl tert-butyl Ether	1.708	1.742	-2.0	89	0.00
20 T	Methylene Chloride	0.595	0.556	6.6	89	0.00
21 T	trans-1,2-Dichloroethene	0.522	0.499	4.4	85	0.00
22 T	Diisopropyl ether	1.763	1.836	-4.1	90	0.00
23 T	Vinyl Acetate	1.556	1.592	-2.3	87	0.00
24 P	1,1-Dichloroethane	0.969	0.983	-1.4	89	0.00
25 T	2-Butanone	0.475	0.446	6.1	82	0.00
26 T	2,2-Dichloropropane	0.793	0.821	-3.5	90	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.614	-1.3	89	0.00
28 T	Bromochloromethane	0.452	0.455	-0.7	91	0.00
29 T	Tetrahydrofuran	0.311	0.288	7.4	80	0.00
30 C	Chloroform	0.992	1.026	-3.4#	92	0.00
31 T	Cyclohexane	0.924	0.871	5.7	86	0.00
32 T	1,1,1-Trichloroethane	0.848	0.888	-4.7	89	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.661	-5.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.313	0.326	-4.2	96	0.00
36 T	1,1-Dichloropropene	0.452	0.455	-0.7	88	0.00
37 T	Ethyl Acetate	0.570	0.541	5.1	84	0.00
38 T	Carbon Tetrachloride	0.460	0.485	-5.4	91	0.00
39 T	Methylcyclohexane	0.571	0.565	1.1	86	0.00
40 TM	Benzene	1.341	1.348	-0.5	87	0.00
41 T	Methacrylonitrile	0.301	0.298	1.0	87	0.00
42 TM	1,2-Dichloroethane	0.496	0.541	-9.1	95	0.00
43 T	Isopropyl Acetate	0.884	0.891	-0.8	86	0.00
44 TM	Trichloroethene	0.368	0.373	-1.4	88	0.00
45 C	1,2-Dichloropropane	0.355	0.364	-2.5#	89	0.00
46 T	Dibromomethane	0.241	0.258	-7.1	92	0.00
47 T	Bromodichloromethane	0.434	0.474	-9.2	92	0.00
48 T	Methyl methacrylate	0.409	0.436	-6.6	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
 Data File : VX023638.D
 Acq On : 10 Aug 2021 17:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 11 03:02:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	84	0.00
50 S	Toluene-d8	1.149	1.219	-6.1	98	0.00
51 T	4-Methyl-2-Pentanone	0.541	0.585	-8.1	90	0.00
52 CM	Toluene	0.831	0.896	-7.8#	92	0.00
53 T	t-1,3-Dichloropropene	0.493	0.555	-12.6	92	0.00
54 T	cis-1,3-Dichloropropene	0.541	0.582	-7.6	89	0.00
55 T	1,1,2-Trichloroethane	0.336	0.381	-13.4	94	0.00
56 T	Ethyl methacrylate	0.539	0.604	-12.1	92	0.00
57 T	1,3-Dichloropropane	0.572	0.625	-9.3	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.286	0.287	-0.3	87	0.00
59 T	2-Hexanone	0.410	0.447	-9.0	91	0.00
60 T	Dibromochloromethane	0.330	0.382	-15.8	93	0.00
61 T	1,2-Dibromoethane	0.359	0.391	-8.9	91	0.00
62 S	4-Bromofluorobenzene	0.391	0.478	-22.3	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.398	0.388	2.5	91	0.00
65 PM	Chlorobenzene	1.023	1.045	-2.2	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.388	-7.5	94	0.00
67 C	Ethyl Benzene	1.775	1.880	-5.9#	96	0.00
68 T	m/p-Xylenes	0.683	0.719	-5.3	96	0.00
69 T	o-Xylene	0.677	0.709	-4.7	97	0.00
70 T	Styrene	1.072	1.201	-12.0	100	0.00
71 P	Bromoform	0.266	0.297	-11.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.773	3.843	-1.9	99	0.00
74 T	N-amyl acetate	1.721	1.732	-0.6	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.229	1.188	3.3	98	0.00
76 T	1,2,3-Trichloropropane	1.100	1.131	-2.8	101	0.00
77 T	Bromobenzene	0.910	0.951	-4.5	102	0.00
78 T	n-propylbenzene	4.270	4.485	-5.0	101	0.00
79 T	2-Chlorotoluene	2.552	2.689	-5.4	104	0.00
80 T	1,3,5-Trimethylbenzene	3.073	3.234	-5.2	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.398	2.7	93	0.00
82 T	4-Chlorotoluene	2.887	3.075	-6.5	104	0.00
83 T	tert-Butylbenzene	3.014	3.201	-6.2	103	0.00
84 T	1,2,4-Trimethylbenzene	3.056	3.232	-5.8	103	0.00
85 T	sec-Butylbenzene	3.908	4.108	-5.1	104	0.00
86 T	p-Isopropyltoluene	3.156	3.410	-8.0	104	0.00
87 T	1,3-Dichlorobenzene	1.684	1.738	-3.2	103	0.00
88 T	1,4-Dichlorobenzene	1.709	1.704	0.3	103	0.00
89 T	n-Butylbenzene	2.857	3.048	-6.7	105	0.00
90 T	Hexachloroethane	0.509	0.549	-7.9	102	0.00
91 T	1,2-Dichlorobenzene	1.630	1.641	-0.7	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.266	0.256	3.8	91	0.00
93 T	1,2,4-Trichlorobenzene	1.024	1.036	-1.2	99	0.00
94 T	Hexachlorobutadiene	0.460	0.449	2.4	101	0.00
95 T	Naphthalene	3.522	3.533	-0.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081021\
Data File : VX023638.D
Acq On : 10 Aug 2021 17:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 11 03:02:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.034	1.053	-1.8	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6