

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
 Data File : VX036912.D
 Acq On : 10 Aug 2023 14:55
 Operator : JC/MD
 Sample : 03960-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 230809056-01-TRIP-BLANK

Quant Time: Aug 11 01:46:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86815	47.163	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.320%
35) Dibromofluoromethane	5.385	113	72869	46.461	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	8.647	98	271774	48.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.079	95	99580	47.455	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.920%
Target Compounds						
16) Acetone	2.380	43	2398	3.139	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
Data File : VX036912.D
Acq On : 10 Aug 2023 14:55
Operator : JC/MD
Sample : 03960-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
230809056-01-TRIP-BLANK

Quant Time: Aug 11 01:46:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

