

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
 Data File : VX036915.D
 Acq On : 10 Aug 2023 16:06
 Operator : JC/MD
 Sample : 03955-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-1

Quant Time: Aug 11 05:15:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215179	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83329	47.951	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.900%
35) Dibromofluoromethane	5.385	113	70133	47.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	8.647	98	259285	49.435	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	99959	50.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.020%
Target Compounds						
16) Acetone	2.386	43	2487	3.449	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
Data File : VX036915.D
Acq On : 10 Aug 2023 16:06
Operator : JC/MD
Sample : 03955-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-1

Quant Time: Aug 11 05:15:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

