

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
 Data File : VX036917.D
 Acq On : 10 Aug 2023 16:55
 Operator : JC/MD
 Sample : 03954-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-16

Quant Time: Aug 11 01:48:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	119707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86421	48.791	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.385	113	71383	46.747	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.500%
50) Toluene-d8	8.647	98	272804	50.378	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.760%
62) 4-Bromofluorobenzene	11.079	95	109627	53.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.320%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14754	20.074	ug/l	91
20) Methylene Chloride	2.788	84	6967	4.262	ug/l #	93
43) Isopropyl Acetate	6.342	43	67776	15.284	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081023\
Data File : VX036917.D
Acq On : 10 Aug 2023 16:55
Operator : JC/MD
Sample : 03954-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-16

Quant Time: Aug 11 01:48:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

