

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081023\  
 Data File : VX036919.D  
 Acq On : 10 Aug 2023 17:42  
 Operator : JC/MD  
 Sample : 03954-11  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 WC-18

Quant Time: Aug 11 01:49:20 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072123W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 21 12:14:12 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 125550   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 232316   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 226963   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 104092   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 89481    | 48.168   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 96.340%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 75371    | 47.201   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 94.400%  |
| 50) Toluene-d8              | 8.647  | 98    | 285253   | 50.374   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 100.740% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 112744   | 52.772   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 105.540% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.386  | 43    | 16514    | 21.422   | ug/l   | 94       |
| 20) Methylene Chloride      | 2.788  | 84    | 12329    | 7.191    | ug/l # | 92       |
| 43) Isopropyl Acetate       | 6.342  | 43    | 93605    | 20.186   | ug/l # | 94       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081023\  
Data File : VX036919.D  
Acq On : 10 Aug 2023 17:42  
Operator : JC/MD  
Sample : 03954-11  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
WC-18

Quant Time: Aug 11 01:49:20 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072123W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 21 12:14:12 2023  
Response via : Initial Calibration

