

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011425.D
 Acq On : 11 Aug 2019 12:45
 Operator : JC/SP
 Sample : PB122149TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB122149TB

Quant Time: Aug 12 08:17:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	212729	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	313619	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	285063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	127022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112449	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
35) Dibromofluoromethane	5.49	113	96064	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	8.71	98	368949	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	11.13	95	124729	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
Target Compounds						
43) Isopropyl Acetate	6.44	43	29011	6.607	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011425.D
Acq On : 11 Aug 2019 12:45
Operator : JC/SP
Sample : PB122149TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB122149TB

Quant Time: Aug 12 08:17:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

