

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011437.D
 Acq On : 11 Aug 2019 17:25
 Operator : JC/SP
 Sample : K4228-36
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2S-I4-(0)

Quant Time: Aug 12 09:06:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	202136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	125685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	108331	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
35) Dibromofluoromethane	5.49	113	91825	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	8.71	98	357385	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
62) 4-Bromofluorobenzene	11.13	95	119377	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
Target Compounds						
16) Acetone	2.43	43	7414	7.652	ug/l	100
20) Methylene Chloride	2.85	84	6508	3.067	ug/l	94
43) Isopropyl Acetate	6.45	43	8305	1.964	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011437.D
Acq On : 11 Aug 2019 17:25
Operator : JC/SP
Sample : K4228-36
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2S-14-(0)

Quant Time: Aug 12 09:06:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

