

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011442.D
 Acq On : 11 Aug 2019 19:22
 Operator : JC/SP
 Sample : K4238-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-18

Quant Time: Aug 12 09:35:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	202762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	302042	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	122073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107346	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	5.49	113	91260	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	8.71	98	360251	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	11.13	95	118488	44.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.18%	
Target Compounds						
16) Acetone	2.43	43	7244	7.453	ug/l	94
20) Methylene Chloride	2.85	84	19340	9.086	ug/l	97
25) 2-Butanone	4.68	43	3473m	2.512	ug/l	
43) Isopropyl Acetate	6.45	43	9120	2.156	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011442.D
Acq On : 11 Aug 2019 19:22
Operator : JC/SP
Sample : K4238-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-18

Quant Time: Aug 12 09:35:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

