

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011454.D
 Acq On : 12 Aug 2019 01:11
 Operator : JC/SP
 Sample : K4266-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R1-R2(A1-A4)-(4.3)

Quant Time: Aug 12 08:34:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	204257	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	304101	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	281604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	125851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110694	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	5.49	113	94411	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	8.71	98	367287	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	11.13	95	120593	45.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.14%	
Target Compounds						
16) Acetone	2.44	43	7239	7.394	ug/l	92
20) Methylene Chloride	2.85	84	44337	20.677	ug/l	98
30) Chloroform	5.21	83	3959	1.134	ug/l #	64
43) Isopropyl Acetate	6.44	43	8499	1.996	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011454.D
Acq On : 12 Aug 2019 01:11
Operator : JC/SP
Sample : K4266-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R1-R2(A1-A4)-(4.3)

Quant Time: Aug 12 08:34:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

