

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011456.D
 Acq On : 12 Aug 2019 01:58
 Operator : JC/SP
 Sample : K4095-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-080919-A

Quant Time: Aug 12 08:49:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	199410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301557	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	275654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	125145	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107793	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
35) Dibromofluoromethane	5.49	113	91651	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
50) Toluene-d8	8.71	98	360108	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.13	95	119973	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
Target Compounds						
16) Acetone	2.43	43	5206	5.446	ug/l	97
20) Methylene Chloride	2.85	84	11602	5.542	ug/l	96
30) Chloroform	5.20	83	3648	1.071	ug/l #	82
43) Isopropyl Acetate	6.44	43	33448	7.922	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011456.D
Acq On : 12 Aug 2019 01:58
Operator : JC/SP
Sample : K4095-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-04-080919-A

Quant Time: Aug 12 08:49:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

