

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011462.D
 Acq On : 12 Aug 2019 04:18
 Operator : JC/SP
 Sample : K4266-24
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R1-R2-C1-C4-(4.3)

Quant Time: Aug 12 09:16:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	195582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	292674	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	123890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107359	54.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.02%	
35) Dibromofluoromethane	5.49	113	90081	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
50) Toluene-d8	8.71	98	354887	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	118757	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
Target Compounds						
16) Acetone	2.43	43	5894	6.287	ug/l	92
20) Methylene Chloride	2.85	84	8659	4.217	ug/l	93
43) Isopropyl Acetate	6.45	43	8370	2.042	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011462.D
Acq On : 12 Aug 2019 04:18
Operator : JC/SP
Sample : K4266-24
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R1-R2-C1-C4-(4.3)

Quant Time: Aug 12 09:16:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

