

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081119\
 Data File : VX011465.D
 Acq On : 12 Aug 2019 05:28
 Operator : JC/SP
 Sample : K4085-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-05-080919-A

Quant Time: Aug 12 09:24:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	198985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	299986	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129627	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107898	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
35) Dibromofluoromethane	5.49	113	91793	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
50) Toluene-d8	8.71	98	362923	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	11.13	95	123205	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
Target Compounds						
16) Acetone	2.43	43	6237	6.539	ug/l	96
20) Methylene Chloride	2.85	84	9941	4.759	ug/l	97
43) Isopropyl Acetate	6.44	43	7341	1.748	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081119\
Data File : VX011465.D
Acq On : 12 Aug 2019 05:28
Operator : JC/SP
Sample : K4085-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
HR-05-080919-A

Quant Time: Aug 12 09:24:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

