


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081119\
Data File : VX011445.D
Acq On : 11 Aug 2019 20:55
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Title : SW846 8260
Last Update : Fri Aug 02 01:55:00 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	6792	PASS
75	95	30	60	51.7	17766	PASS
95	95	100	100	100.0	34392	PASS
96	95	5	9	6.3	2155	PASS
173	174	0.00	2	1.7	507	PASS
174	95	50	100	88.5	30445	PASS
175	174	5	9	8.0	2433	PASS
176	174	95	101	98.3	29928	PASS
177	176	5	9	7.1	2118	PASS