

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081121\
 Data File : VX023648.D
 Acq On : 11 Aug 2021 13:22
 Operator : JC/MD
 Sample : VX0811WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK605

Quant Time: Aug 12 01:27:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML081121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 11 12:02:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	328334	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	297103	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	131646	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	86033	50.271	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.540%
7) Chloroethane-d5	1.666	69	75086	53.409	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.820%
11) 1,1-Dichloroethene-d2	2.306	63	148464	39.938	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.880%
21) 2-Butanone-d5	4.465	46	183569	111.259	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.260%
24) Chloroform-d	5.068	84	195106	50.348	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.700%
26) 1,2-Dichloroethane-d4	5.964	65	137475	51.752	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.500%
32) Benzene-d6	5.983	84	408287	54.553	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.100%
36) 1,2-Dichloropropane-d6	7.312	67	129017	54.787	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.580%
41) Toluene-d8	8.653	98	378524	52.515	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.040%
43) trans-1,3-Dichloroprop...	8.952	79	55647	49.935	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.880%
47) 2-Hexanone-d5	9.384	63	132169	116.742	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	116.740%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	165936	52.896	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	12.323	152	127201	55.062	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081121\
Data File : VX023648.D
Acq On : 11 Aug 2021 13:22
Operator : JC/MD
Sample : VX0811WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK605

Quant Time: Aug 12 01:27:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML081121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 11 12:02:06 2021
Response via : Initial Calibration

