

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081121\
 Data File : VX023654.D
 Acq On : 11 Aug 2021 17:16
 Operator : JC/MD
 Sample : M3364-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DBK84

Quant Time: Aug 12 01:29:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML081121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 11 12:02:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	317650	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	293238	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	128582	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	74769	45.159	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.320%
7) Chloroethane-d5	1.666	69	65977	48.508	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.020%
11) 1,1-Dichloroethene-d2	2.306	63	128593	35.756	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.520%
21) 2-Butanone-d5	4.465	46	157552	98.702	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.700%
24) Chloroform-d	5.068	84	175134	46.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.420%
26) 1,2-Dichloroethane-d4	5.964	65	124110	48.292	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.580%
32) Benzene-d6	5.983	84	358255	48.499	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	7.318	67	114310	49.182	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.360%
41) Toluene-d8	8.653	98	333335	46.855	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.720%
43) trans-1,3-Dichloroprop...	8.952	79	47246	42.955	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.920%
47) 2-Hexanone-d5	9.391	63	113092	101.208	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.210%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	153295	49.511	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.020%
66) 1,2-Dichlorobenzene-d4	12.323	152	115932	51.380	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.760%
Target Compounds						Qvalue
13) Acetone	2.386	43	5644	4.016	ug/L	90
18) Methyl tert-butyl Ether	3.129	73	5090	0.803	ug/L #	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081121\
Data File : VX023654.D
Acq On : 11 Aug 2021 17:16
Operator : JC/MD
Sample : M3364-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DBK84

Quant Time: Aug 12 01:29:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML081121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 11 12:02:06 2021
Response via : Initial Calibration









#13
Acetone
Concen: 4.016 ug/L
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX023654.D
Acq: 11 Aug 2021 17:16

Instrument :
MSVOA_X
ClientSampled :
DBK84

Tgt Ion: 43 Resp: 5644
Ion Ratio Lower Upper
43 100
58 28.3 0.0 68.2



#18
Methyl tert-butyl Ether
Concen: 0.803 ug/L
RT: 3.129 min Scan# 335
Delta R.T. 0.006 min
Lab File: VX023654.D
Acq: 11 Aug 2021 17:16

Tgt Ion: 73 Resp: 5090
Ion Ratio Lower Upper
73 100
43 26.5 15.1 22.7#
57 20.5 17.4 26.2

