

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081122\
 Data File : VX030572.D
 Acq On : 11 Aug 2022 14:43
 Operator : JC/MD
 Sample : VIBLK670
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK670

Quant Time: Aug 12 04:19:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 12 04:17:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	265976	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	242474	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	105836	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	78450	45.930	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.860%
7) Chloroethane-d5	1.666	69	65990	54.805	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.600%
11) 1,1-Dichloroethene-d2	2.307	63	125350	38.797	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.600%
21) 2-Butanone-d5	4.477	46	197512	124.802	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.800%
24) Chloroform-d	5.068	84	192957	54.878	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.760%
26) 1,2-Dichloroethane-d4	5.964	65	139210	58.310	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.620%
32) Benzene-d6	5.977	84	361031	51.980	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.960%
36) 1,2-Dichloropropane-d6	7.318	67	121323	53.839	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.680%
41) Toluene-d8	8.653	98	326716	49.903	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.800%
43) trans-1,3-Dichloroprop...	8.952	79	55410	49.427	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.860%
47) 2-Hexanone-d5	9.391	63	129416	112.670	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.670%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	167165	56.469	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.940%
66) 1,2-Dichlorobenzene-d4	12.323	152	116373	56.690	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081122\
Data File : VX030572.D
Acq On : 11 Aug 2022 14:43
Operator : JC/MD
Sample : VIBLK670
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK670

Quant Time: Aug 12 04:19:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 12 04:17:17 2022
Response via : Initial Calibration

