

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
 Data File : VX036938.D
 Acq On : 11 Aug 2023 14:03
 Operator : JC/MD
 Sample : 03980-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 11 22:58:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118541	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84401	48.119	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.240%
35) Dibromofluoromethane	5.391	113	70627	47.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	8.647	98	262249	49.547	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	100592	50.374	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
Data File : VX036938.D
Acq On : 11 Aug 2023 14:03
Operator : JC/MD
Sample : 03980-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 11 22:58:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

