

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
 Data File : VX036940.D
 Acq On : 11 Aug 2023 14:51
 Operator : JC/MD
 Sample : 03955-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-26

Quant Time: Aug 11 22:59:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210456	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88112	48.613	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.220%
35) Dibromofluoromethane	5.385	113	71578	46.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.580%
50) Toluene-d8	8.647	98	271977	49.600	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.079	95	102228	49.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.820%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	21104	28.059	ug/l	92
20) Methylene Chloride	2.788	84	18751	11.210	ug/l	90
25) 2-Butanone	4.568	43	7420	6.081	ug/l	96
43) Isopropyl Acetate	6.342	43	105965	23.599	ug/l #	94
95) Naphthalene	13.774	128	160918	24.803	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
Data File : VX036940.D
Acq On : 11 Aug 2023 14:51
Operator : JC/MD
Sample : 03955-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-26

Quant Time: Aug 11 22:59:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

