

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
 Data File : VX036941.D
 Acq On : 11 Aug 2023 15:15
 Operator : JC/MD
 Sample : 03955-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-27

Quant Time: Aug 11 22:59:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	242010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94064	48.753	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.500%
35) Dibromofluoromethane	5.385	113	77352	46.501	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.647	98	293325	49.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.460%
62) 4-Bromofluorobenzene	11.079	95	115066	51.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.400%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18263	22.811	ug/l	94
20) Methylene Chloride	2.788	84	14562	8.178	ug/l	89
37) Ethyl Acetate	4.733	43	3084	1.099	ug/l #	89
43) Isopropyl Acetate	6.342	43	100171	20.737	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
Data File : VX036941.D
Acq On : 11 Aug 2023 15:15
Operator : JC/MD
Sample : 03955-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-27

Quant Time: Aug 11 22:59:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

