

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
 Data File : VX036951.D
 Acq On : 11 Aug 2023 19:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 11 23:06:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93792	47.174	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.340%
35) Dibromofluoromethane	5.385	113	80061	46.695	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	8.647	98	299956	49.333	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	122666	53.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.940%
Target Compounds						
16) Acetone	2.386	43	1860	2.254	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
Data File : VX036951.D
Acq On : 11 Aug 2023 19:15
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 11 23:06:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

