

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
 Data File : VX036937.D
 Acq On : 11 Aug 2023 13:39
 Operator : JC/MD
 Sample : 03980-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 11 22:58:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86414	47.615	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.220%
35) Dibromofluoromethane	5.385	113	70927	46.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.240%
50) Toluene-d8	8.647	98	269566	49.428	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.079	95	103677	50.388	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.780%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081123\
Data File : VX036937.D
Acq On : 11 Aug 2023 13:39
Operator : JC/MD
Sample : 03980-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 11 22:58:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 122655
Ion Ratio Lower Upper
168 100
99 60.4 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 47.615 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39

Tgt Ion: 65 Resp: 86414
Ion Ratio Lower Upper
65 100
67 50.9 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 223744
Ion Ratio Lower Upper
114 100
63 20.5 0.0 45.0
88 16.8 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.120 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39

Tgt Ion:113 Resp: 70927
Ion Ratio Lower Upper
113 100
111 103.8 82.9 124.3
192 18.3 13.3 19.9









#50
Toluene-d8
Concen: 49.428 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 269566
Ion Ratio Lower Upper
98 100
100 65.8 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 50.388 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39

Tgt Ion: 95 Resp: 103677
Ion Ratio Lower Upper
95 100
174 71.3 0.0 135.6
176 69.4 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 208685
Ion Ratio Lower Upper
117 100
82 56.7 46.2 69.4
119 32.5 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX036937.D
Acq: 11 Aug 2023 13:39

Tgt Ion:152 Resp: 91133
Ion Ratio Lower Upper
152 100
115 63.0 43.3 129.9
150 158.6 0.0 345.0

