

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011484.D
 Acq On : 12 Aug 2019 15:29
 Operator : JC/SP
 Sample : K4278-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0705

Quant Time: Aug 13 08:23:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	280746	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114367	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	102097	54.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.78%	
35) Dibromofluoromethane	5.49	113	85966	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	8.71	98	328517	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
62) 4-Bromofluorobenzene	11.13	95	108680	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	
Target Compounds						
16) Acetone	2.43	43	4716	5.278	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011484.D
Acq On : 12 Aug 2019 15:29
Operator : JC/SP
Sample : K4278-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0705

Quant Time: Aug 13 08:23:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

