

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011491.D
 Acq On : 12 Aug 2019 18:13
 Operator : JC/SP
 Sample : K4278-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0691

Quant Time: Aug 13 08:39:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	183921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	273786	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	111148	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	101776	54.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.90%	
35) Dibromofluoromethane	5.49	113	84904	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
50) Toluene-d8	8.71	98	321243	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
62) 4-Bromofluorobenzene	11.13	95	105567	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
Target Compounds						
4) Vinyl Chloride	1.40	62	12590	7.175	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	3415	1.929	ug/l #	83
27) cis-1,2-Dichloroethene	4.59	96	53262	26.123	ug/l	94
40) Benzene	6.14	78	27553	3.863	ug/l	94
65) Chlorobenzene	10.13	112	288281	54.897	ug/l	99
87) 1,3-Dichlorobenzene	12.02	146	1107	0.297	ug/l	86
88) 1,4-Dichlorobenzene	12.10	146	10038	2.614	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011491.D
Acq On : 12 Aug 2019 18:13
Operator : JC/SP
Sample : K4278-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0691

Quant Time: Aug 13 08:39:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

