

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011506.D
 Acq On : 13 Aug 2019 00:49
 Operator : JC/SP
 Sample : K4230-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-3A-(2)

Quant Time: Aug 13 09:17:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	191034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	288935	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	267879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116534	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	106670	55.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.90%	
35) Dibromofluoromethane	5.49	113	88766	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
50) Toluene-d8	8.71	98	346392	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.13	95	112202	44.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.28%	
Target Compounds						
20) Methylene Chloride	2.85	84	2309	1.151	ug/l	90
43) Isopropyl Acetate	6.45	43	20746	5.128	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011506.D
Acq On : 13 Aug 2019 00:49
Operator : JC/SP
Sample : K4230-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-3A-(2)

Quant Time: Aug 13 09:17:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

