

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011508.D
 Acq On : 13 Aug 2019 01:36
 Operator : JC/SP
 Sample : K4230-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-3E-(2)

Quant Time: Aug 13 09:20:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	183744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	278319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	112109	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	103365	55.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.72%	
35) Dibromofluoromethane	5.49	113	87265	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	8.71	98	335212	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	11.13	95	109095	44.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.10%	
Target Compounds						
20) Methylene Chloride	2.85	84	2107	1.092	ug/l #	85
43) Isopropyl Acetate	6.44	43	19565	5.021	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011508.D
Acq On : 13 Aug 2019 01:36
Operator : JC/SP
Sample : K4230-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-3E-(2)

Quant Time: Aug 13 09:20:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

