

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011514.D
 Acq On : 13 Aug 2019 03:57
 Operator : JC/SP
 Sample : K4278-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0709

Quant Time: Aug 13 09:32:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	179073	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	269847	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	239426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	107379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	101279	56.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.32%	
35) Dibromofluoromethane	5.49	113	84472	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
50) Toluene-d8	8.71	98	317578	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	11.13	95	103907	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1659	0.971	ug/l	93
21) trans-1,2-Dichloroethene	3.16	96	2648	1.536	ug/l #	81
27) cis-1,2-Dichloroethene	4.59	96	41719	21.015	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011514.D
Acq On : 13 Aug 2019 03:57
Operator : JC/SP
Sample : K4278-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0709

Quant Time: Aug 13 09:32:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

