

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011515.D
 Acq On : 13 Aug 2019 04:20
 Operator : JC/SP
 Sample : K4278-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0706

Quant Time: Aug 13 09:34:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	180223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	274238	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	239622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	106190	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	102170	56.29	ug/l	0.00
Spiked Amount						
			Recovery	=	112.58%	
35) Dibromofluoromethane	5.49	113	85304	47.64	ug/l	0.00
Spiked Amount						
			Recovery	=	95.28%	
50) Toluene-d8	8.71	98	318998	47.20	ug/l	0.00
Spiked Amount						
			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	11.13	95	101979	42.26	ug/l	0.00
Spiked Amount						
			Recovery	=	84.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011515.D
Acq On : 13 Aug 2019 04:20
Operator : JC/SP
Sample : K4278-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0706

Quant Time: Aug 13 09:34:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

