

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081219\
 Data File : VX011517.D
 Acq On : 13 Aug 2019 05:07
 Operator : JC/SP
 Sample : K4278-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0696

Quant Time: Aug 13 09:38:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	187483	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	284190	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110467	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	106128	56.21	ug/l	0.00
Spiked Amount			Recovery	=	112.42%	
35) Dibromofluoromethane	5.49	113	88774	47.85	ug/l	0.00
Spiked Amount			Recovery	=	95.70%	
50) Toluene-d8	8.71	98	332209	47.43	ug/l	0.00
Spiked Amount			Recovery	=	94.86%	
62) 4-Bromofluorobenzene	11.13	95	108244	43.29	ug/l	0.00
Spiked Amount			Recovery	=	86.58%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081219\
Data File : VX011517.D
Acq On : 13 Aug 2019 05:07
Operator : JC/SP
Sample : K4278-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0696

Quant Time: Aug 13 09:38:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

