

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081220\
 Data File : VX017857.D
 Acq On : 12 Aug 2020 13:11
 Operator : JC/SP
 Sample : VX0812WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK65

Quant Time: Aug 13 01:26:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 13 01:19:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.84	114	324943	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	293930	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	138992	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	107991	4.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.20%
7) Chloroethane-d5	1.69	69	80308	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	2.34	63	155337	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.54	46	261590	43.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.74%
24) Chloroform-d	5.15	84	184188	4.04	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
26) 1,2-Dichloroethane-d4	6.03	65	106711	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
32) Benzene-d6	6.05	84	369065	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	7.37	67	98522	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.00%
41) Toluene-d8	8.70	98	314777	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
45) trans-1,3-Dichloropropene-	8.99	79	44207	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
48) 2-Hexanone-d5	9.43	63	174224	38.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.62%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	74828	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.00%
65) 1,2-Dichlorobenzene-d4	12.36	152	110117	4.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081220\
Data File : VX017857.D
Acq On : 12 Aug 2020 13:11
Operator : JC/SP
Sample : VX0812WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK65

Quant Time: Aug 13 01:26:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 13 01:19:50 2020
Response via : Initial Calibration

