

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081318\
 Data File : VX004005.D
 Acq On : 13 Aug 2018 17:28
 Operator : JC/MD
 Sample : J4447-45
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-1-10

Quant Time: Aug 14 05:16:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 13 01:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	304798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	512780	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	477301	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	261332	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	228257	56.49	ug/l	0.00
Spiked Amount			Recovery	=	112.98%	
35) Dibromofluoromethane	5.51	113	183074	47.13	ug/l	0.00
Spiked Amount			Recovery	=	94.26%	
50) Toluene-d8	8.72	98	731230	48.40	ug/l	0.00
Spiked Amount			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	11.14	95	253519	51.86	ug/l	0.00
Spiked Amount			Recovery	=	103.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	14948	6.22	ug/l	97
18) Methyl Acetate	2.78	43	6242	1.19	ug/l	96
20) Methylene Chloride	2.86	84	230522	39.56	ug/l	98
25) 2-Butanone	4.70	43	3668	1.01	ug/l #	84
30) Chloroform	5.22	83	6592	1.02	ug/l	96
43) Isopropyl Acetate	6.46	43	156846	16.50	ug/l	99
68) m/p-Xylenes	10.36	106	8119	1.02	ug/l	100
74) N-amyl acetate	6.46	43	156846	15.60	ug/l #	99
95) Naphthalene	13.83	128	96044	4.93	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081318\
Data File : VX004005.D
Acq On : 13 Aug 2018 17:28
Operator : JC/MD
Sample : J4447-45
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-1-10

Quant Time: Aug 14 05:16:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 13 01:34:27 2018
Response via : Initial Calibration

