

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081318\
 Data File : VX004009.D
 Acq On : 13 Aug 2018 19:14
 Operator : JC/MD
 Sample : J4442-30
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-1-3

Quant Time: Aug 14 05:24:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 13 01:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	310716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	483359	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	437028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	237726	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	216855	52.65	ug/l	0.00
Spiked Amount			Recovery	=	105.30%	
35) Dibromofluoromethane	5.51	113	175225	47.85	ug/l	0.00
Spiked Amount			Recovery	=	95.70%	
50) Toluene-d8	8.72	98	675104	47.41	ug/l	0.00
Spiked Amount			Recovery	=	94.82%	
62) 4-Bromofluorobenzene	11.14	95	229610	49.83	ug/l	0.00
Spiked Amount			Recovery	=	99.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	15640	6.38	ug/l	99
18) Methyl Acetate	2.77	43	6425	1.20	ug/l	97
20) Methylene Chloride	2.86	84	10163	1.71	ug/l	95
25) 2-Butanone	4.70	43	3870	1.05	ug/l #	89
30) Chloroform	5.21	83	13995	2.12	ug/l	99
43) Isopropyl Acetate	6.46	43	167633	18.71	ug/l	100
74) N-amyl acetate	6.46	43	167633	18.33	ug/l #	100
95) Naphthalene	13.83	128	303254	17.13	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081318\
Data File : VX004009.D
Acq On : 13 Aug 2018 19:14
Operator : JC/MD
Sample : J4442-30
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-1-3

Quant Time: Aug 14 05:24:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 13 01:34:27 2018
Response via : Initial Calibration

