

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081318\
 Data File : VX004012.D
 Acq On : 13 Aug 2018 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 05:38:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 13 01:34:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.507	0.488	3.7	83	0.00
3 P	Chloromethane	0.689	0.589	14.5	79	0.00
4 C	Vinyl Chloride	0.688	0.655	4.8#	86	0.00
5 T	Bromomethane	0.438	0.328	25.1#	74	0.00
6 T	Chloroethane	0.440	0.445	-1.1	86	0.00
7 T	Trichlorofluoromethane	0.908	0.909	-0.1	91	0.00
8 T	Diethyl Ether	0.408	0.419	-2.7	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.559	1.4	92	0.00
10 T	Methyl Iodide	0.706	0.591	16.3	72	0.00
11 T	Tert butyl alcohol	0.159	0.158	0.6	89	-0.03
12 CM	1,1-Dichloroethene	0.537	0.534	0.6#	91	0.00
13 T	Acrolein	0.088	0.066	25.0#	73	0.00
14 T	Allyl chloride	1.035	1.001	3.3	86	0.00
15 T	Acrylonitrile	0.380	0.373	1.8	89	-0.01
16 T	Acetone	0.394	0.371	5.8	87	0.00
17 T	Carbon Disulfide	1.637	1.379	15.8	82	0.01
18 T	Methyl Acetate	0.859	0.927	-7.9	98	-0.01
19 T	Methyl tert-butyl Ether	1.902	2.027	-6.6	96	-0.01
20 T	Methylene Chloride	0.956	0.642	32.8#	91	0.00
21 T	trans-1,2-Dichloroethene	0.609	0.597	2.0	92	0.00
22 T	Diisopropyl ether	1.902	2.065	-8.6	99	0.00
23 T	Vinyl Acetate	1.582	1.718	-8.6	96	0.00
24 P	1,1-Dichloroethane	1.130	1.137	-0.6	91	0.00
25 T	2-Butanone	0.593	0.627	-5.7	93	-0.02
26 T	2,2-Dichloropropane	0.815	0.757	7.1	85	0.00
27 T	cis-1,2-Dichloroethene	0.661	0.694	-5.0	97	0.00
28 T	Bromochloromethane	0.483	0.488	-1.0	90	0.00
29 T	Tetrahydrofuran	0.305	0.319	-4.6	93	-0.01
30 C	Chloroform	1.063	1.122	-5.6#	95	0.00
31 T	Cyclohexane	0.955	0.974	-2.0	91	0.00
32 T	1,1,1-Trichloroethane	0.892	0.956	-7.2	95	0.00
33 S	1,2-Dichloroethane-d4	0.663	0.680	-2.6	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.379	0.359	5.3	87	0.00
36 T	1,1-Dichloropropene	0.560	0.548	2.1	93	0.00
37 T	Ethyl Acetate	0.616	0.625	-1.5	95	-0.02
38 T	Carbon Tetrachloride	0.533	0.541	-1.5	94	0.00
39 T	Methylcyclohexane	0.660	0.646	2.1	92	0.00
40 TM	Benzene	1.682	1.702	-1.2	96	0.00
41 T	Methacrylonitrile	0.339	0.353	-4.1	97	-0.01
42 TM	1,2-Dichloroethane	0.569	0.587	-3.2	98	0.00
43 T	Isopropyl Acetate	0.927	0.978	-5.5	97	0.00
44 TM	Trichloroethene	0.460	0.461	-0.2	98	0.00
45 C	1,2-Dichloropropane	0.434	0.444	-2.3#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081318\
 Data File : VX004012.D
 Acq On : 13 Aug 2018 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 05:38:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 13 01:34:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.282	0.288	-2.1	97	0.00
47 T	Bromodichloromethane	0.531	0.559	-5.3	97	0.00
48 T	Methyl methacrylate	0.471	0.506	-7.4	96	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	91	0.00
50 S	Toluene-d8	1.473	1.429	3.0	88	0.00
51 T	4-Methyl-2-Pentanone	0.768	0.811	-5.6	95	0.00
52 CM	Toluene	1.063	1.102	-3.7#	95	0.00
53 T	t-1,3-Dichloropropene	0.600	0.634	-5.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.647	0.683	-5.6	94	0.00
55 T	1,1,2-Trichloroethane	0.434	0.455	-4.8	98	0.00
56 T	Ethyl methacrylate	0.630	0.698	-10.8	98	0.00
57 T	1,3-Dichloropropane	0.727	0.766	-5.4	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.310	0.318	-2.6	90	0.00
59 T	2-Hexanone	0.574	0.619	-7.8	96	0.00
60 T	Dibromochloromethane	0.405	0.462	-14.1	102	0.00
61 T	1,2-Dibromoethane	0.435	0.472	-8.5	101	0.00
62 S	4-Bromofluorobenzene	0.477	0.540	-13.2	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.492	0.474	3.7	100	0.00
65 PM	Chlorobenzene	1.301	1.308	-0.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.447	0.475	-6.3	100	0.00
67 C	Ethyl Benzene	2.130	2.213	-3.9#	100	0.00
68 T	m/p-Xylenes	0.832	0.891	-7.1	103	0.00
69 T	o-Xylene	0.786	0.856	-8.9	102	0.00
70 T	Styrene	1.286	1.449	-12.7	106	0.00
71 P	Bromoform	0.334	0.387	-15.9	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.699	3.646	1.4	102	0.00
74 T	N-amyl acetate	1.923	1.684	12.4	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.268	1.225	3.4	114	0.00
76 T	1,2,3-Trichloropropane	1.104	1.083	1.9	112	0.00
77 T	Bromobenzene	0.987	0.979	0.8	116	0.00
78 T	n-propylbenzene	4.153	4.193	-1.0	112	0.00
79 T	2-Chlorotoluene	2.532	2.465	2.6	112	0.00
80 T	1,3,5-Trimethylbenzene	3.016	3.036	-0.7	112	0.00
81 T	trans-1,4-Dichloro-2-butene	0.362	0.345	4.7	101	0.00
82 T	4-Chlorotoluene	2.967	2.881	2.9	111	0.00
83 T	tert-Butylbenzene	2.967	3.099	-4.4	114	0.00
84 T	1,2,4-Trimethylbenzene	3.033	3.131	-3.2	112	0.00
85 T	sec-Butylbenzene	3.502	3.632	-3.7	111	0.00
86 T	p-Isopropyltoluene	3.142	3.285	-4.6	112	0.00
87 T	1,3-Dichlorobenzene	1.824	1.808	0.9	115	0.00
88 T	1,4-Dichlorobenzene	1.892	1.836	3.0	115	0.00
89 T	n-Butylbenzene	2.738	2.559	6.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081318\
Data File : VX004012.D
Acq On : 13 Aug 2018 20:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 05:38:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 13 01:34:27 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.510	0.454	11.0	96	0.00
91 T	1,2-Dichlorobenzene	1.808	1.693	6.4	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.260	0.225	13.5	93	0.00
93 T	1,2,4-Trichlorobenzene	1.300	1.216	6.5	105	0.00
94 T	Hexachlorobutadiene	0.609	0.555	8.9	108	0.00
95 T	Naphthalene	3.724	3.962	-6.4	109	0.00
96 T	1,2,3-Trichlorobenzene	1.298	1.327	-2.2	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6