

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081319\
 Data File : VX011527.D
 Acq On : 13 Aug 2019 11:47
 Operator : JC/SP
 Sample : K4279-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB-2019-1

Quant Time: Aug 14 01:25:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	194957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	294025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114203	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	109648	55.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.70%	
35) Dibromofluoromethane	5.49	113	90258	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
50) Toluene-d8	8.71	98	344285	47.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.02%	
62) 4-Bromofluorobenzene	11.13	95	113882	44.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081319\
Data File : VX011527.D
Acq On : 13 Aug 2019 11:47
Operator : JC/SP
Sample : K4279-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB-2019-1

Quant Time: Aug 14 01:25:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

