

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081319\
 Data File : VX011539.D
 Acq On : 13 Aug 2019 16:27
 Operator : JC/SP
 Sample : K4279-04DL 40X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0670DL

Quant Time: Aug 14 11:37:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	196832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	294387	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	263181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	121401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	108826	54.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.80%	
35) Dibromofluoromethane	5.49	113	91952	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	8.71	98	341061	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
62) 4-Bromofluorobenzene	11.13	95	118277	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
Target Compounds						
40) Benzene	6.14	78	278357	36.292	ug/l	97
67) Ethyl Benzene	10.25	91	72489	7.670	ug/l	98
68) m/p-Xylenes	10.36	106	3920	1.072	ug/l	98
73) Isopropylbenzene	11.02	105	8795	1.059	ug/l	96
78) n-propylbenzene	11.36	91	10454	1.132	ug/l	100
95) Naphthalene	13.83	128	21493	2.711	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081319\
Data File : VX011539.D
Acq On : 13 Aug 2019 16:27
Operator : JC/SP
Sample : K4279-04DL 40X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0670DL

Quant Time: Aug 14 11:37:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

